



Statement of Memory Volatility

Product and Manufacturer				
Models: 393/393 FC CAT III 1500V TRMS Clamp Meter		Manufacturer: Fluke Corporation		
		Address: 6920 Seaway Blvd.		
		City: Everett	State: WA	Zip: 98203
Volatile Memory				
Does the item contain volatile memory (i.e., memory whose contents are lost when power is removed)? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No (If the answer is 'Yes', provide the following information for each type of memory.)				
Memory Type (SRAM, DRAM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
MSP430F47187 Internal SRAM	8 KB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Measurement and control	Power off
CC2541F256 Internal SRAM	8 KB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Communication and control	Power off
Non-Volatile Memory				
Does the item contain non-volatile memory (i.e., memory whose contents are retained when power is removed)? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No (If the answer is 'Yes', provide the following information for each type of memory.)				

Memory Type (BBRAM, Flash, EEPROM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:	
MSP430F47187 Internal Flash	112 kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Measurement and control	NA	
CC2541F256 Internal Flash	256 kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Communication and control	NA	
AT25DF081A Internal Flash	1 MB	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Test result storage	Clear with keypad or APP	
Media					
Does the item contain media storage capability (i.e., removable or non-removable disk drives, memory cards, etc.)?					
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No (If the answer is 'Yes', provide the following information for each type of memory.)					
	Storage Type (Disk, Tape, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
			<input type="checkbox"/> Yes <input type="checkbox"/> No		
Additional Information					
NA					
Fluke Representative Information					
Name: Susan Isaac	Title: Product Manager	Office Phone: 1-425-347-6100	Fax/Email: susan.isaac@fluke.com		